

USP <797> Chapter 2.3 Competency Testing in Aseptic Manipulation

For additional Cultivate™ Directions for Use and test logs go to www.parasolmed.com/cultivate

Definitions

Negative: Light to medium amber, clear, no precipitates

Positive: Characterized by turbidity, haziness, specks, or flocculation in the medium.

FV100 - FTM Growth Media

NEGATIVE



POSITIVE



Escherichia coli



Staphylococcus aureus

FV020 - FTM Growth Media

NEGATIVE



POSITIVE



Escherichia coli



Staphylococcus aureus