[1]U.S. Pharmacopeia Convention. (2022). <797> Pharmaceutical Compounding–Sterile Preparations.

© Copyright 2023 Parasol Medical, LLC, all rights reserved 2023_Contact_TD100_SS_Log_REV_A

Cultivate
A PARASOL MEDICAL PRODUCT

Parasol Medical LLC.
1000 Asbury Dr. Suite 1, Buffalo Grove, 60089
E: Customerserivce@parasolmed.com | P: 847.520.3000

Contact™ Surface Sampling, *One Sample per Location*

Cat. #: TD100

2023 USP <797> Section 6.3 Monitoring Surfaces for Viable Particles

For additional Cultivate™ directions for use and test logs, go to www.parasolmed.com/cultivate/cultivate-resources

Person(s) Collecting Samples : _____

Person(s) Documenting Results: _____

Person(s) Identifying / Reading Results: _____

Test Type / Description: _____

Additional Information: _____

PRODUCT INFORMATION

Component No.: T-D100
Media Type: TSA with L&T, Sterile
Media Manufacture: Parasol Medical

Media Lot. #: _____
List all if using multiple lots, identify which Sample No. are associated with each lot

Media Exp.Date: _____
List all if using multiple lots, identify which Sample No. are associated with each lot

☐ *Check box to confirm you collected certificate of analysis (CoA) for growth media.Keep a copy of the CoA(s) on file. To download a copy, visit www.parasolmed.com/cultivate/cultivate-resources*

				INTERVAL 1 2023 USP <797> 30-35°C for no less than 48 hr			INTERVAL 2 2023 USP <797> 20-25°C for no less than 5 days			RESULTS	
Sample No.	Sample Date	Sample Location Decription **Must be described in facility SOPs**	Sample Location ISO Class	Interval Start Date	Growth After 1st Interval (Y or N)	CFU Count	Interval Start Date	Growth After 1st Interval (Y or N)	CFU Count	Test Results	Genus Classification **if CFU count exceeds action levels**
										☐ PASS ☐ FAIL	
										☐ PASS ☐ FAIL	
										☐ PASS ☐ FAIL	
										☐ PASS ☐ FAIL	
										☐ PASS ☐ FAIL	
										☐ PASS ☐ FAIL	
										☐ PASS ☐ FAIL	
										☐ PASS ☐ FAIL	
										☐ PASS ☐ FAIL	
										☐ PASS ☐ FAIL	

CORRECTIVE ACTION SUMMARY / COMMENTS:

Surface Sampling Action Levels, per 2023 USP <797>[1]

ISO Class	Surface Sampling Action Levels (cfu/device or swab)
5	>3
7	>5
8	>50

Action levels are based on the total cfu count for both sides of paddle